

Entered 15/2/17 

96-19784

West Gippsland Waterwatch Volunteer Water Quality Monitoring Data Sheet

RECEIVED
27 MAR 2016
BY:

Site Details

Site Code: TYR 090

Site Name:

Site Grid Reference Easting:
Northing:

Test Details

Date: 20/3/16

Time: 0930 AM PM

Person(s) Conducting the Survey / Test: LEANNE STUMBERT
PETER MAFFEI

Name of Monitoring Group: FRIENDS OF TYROS PARK.

Your Meter - Reading Before Calibration (when in CAL Mode)				Your Meter - Reading Post Calibration (after pressing the ENT button)			
EC Meter	1400			EC Meter	1410		
pH Meter in 7.01	7.6	pH Meter in 4.01	4.0	pH Meter in 7.01	7.0	pH Meter in 4.01	4.0

Water Condition (Circle)

Steady Flow Rising Peak Falling Stagnant (pools) Dry (no water)

Test Results

Air Temperature	10.7	°C
Water Temperature	17.3	°C
pH	5.7	
Electrical Conductivity	80	<input checked="" type="checkbox"/> µS <input type="checkbox"/> mS
Reactive Phosphorus	0.03	mg/L P
Turbidity	7	N.T.U.
Dissolved Oxygen:		<input type="checkbox"/> mg/L <input type="checkbox"/> %

Any Comments for Your Facilitator?

Weather Conditions at Time of Sampling

Sunny Cloudy

Overcast Raining

Windy

Water Appearance

Clear Muddy

Milky Foamy / Frothy

Stained Green Stained Brown

Smelly Scummy

Oily Other:

Last Rainfall


More than a week ago During the last week

During the last 24 hours Raining Now

Amount of rain: 25 mm

Drains

Drain Flow: Drain flowing Drain not flowing



Post: West Gippsland Waterwatch
Reply Paid 1374 Traralgon 3844 or Fax: 51757899